



Fraunhofer

TESTED[®] DEVICE

REGIOLUX GmbH

VSG glass pane

Report No. RE 2011-1189

Single product
Chemical
Resistance

Qualification Certificate

This is to certify that the product mentioned above, provided by

Regiolux GmbH
Königsberg, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number RE 2011-1189.

The chemical resistance of the RX VSG glass pane (color: transparent) was classified according to ISO 4628-1 and VDI 2083 Part 17 with the following result:

Chemical resistance (24 h)			
Formalin 37 %	excellent	Peracetic acid 15 %	excellent
Ammoniac 25 %	excellent	Isopropanol 100 %	good
Hydrogen peroxide 30 %	excellent	Sodium hydroxide 5 %	excellent
Sulfuric acid 5 %	excellent	Sodium hypochlorite 5 %	excellent

DUPLICAT

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under www.tested-device.com.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

RE 2011-1189
Report No. first document

Stuttgart, April 19, 2021
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA